

**Premarket Notification 510(k)**

**Auropol 1**

**5. 510 (k) Summary**

K012839

DEC 31 2001

**Submitter of 510(k):** Wieland Edelmetalle GmbH & Co.  
Schwenninger Str. 13  
D-75179 Pforzheim  
Germany  
Phone: +49-7231-3705-0

**Contact person:** Dr. Gerhard Polzer  
**Phone:** +49-7231-3705-219  
**Fax:** +49-7231-357959  
**e-mail:** gerhard.polzer@wieland-dental.de

**Date of Summary:** 2001-07-21

**Trade name:** Auropol 1

**Classification name:** Alloy, gold based, for clinical use  
**Product code:** EJT  
**C.D.R section:** 872.3060  
**Classification:** Class II

**Legally marketed  
equivalent device:** Pontor 2

**510(k) number:** K911542

**Device description**

Auropol 1 is an extra-hard gold-silver casting alloy (67% noble metals), intended for dental technicians to fabricate dental restorations.

It has an indication which ranges from inlays/onlays and crowns up to long span bridges with two or more pontics and to removable partials. In addition, it is suitable for telescopic and milling work.

Auropol 1 is highly corrosion resistant and has an excellent biocompatibility. It fully complies to the international standard ISO 8891 and fulfills the essential requirements of the European directive 93/42/ECC concerning medical devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

DEC 31 2001

Mr. Gerhard Polzer  
Director, Regulatory Affairs  
Wieland Edelmetalle GmbH & Company  
Schwenninger Strabe 13  
D-75179 Pforzheim  
GERMANY

Re: K012839

Trade/Device Name: Auropal I  
Regulation Number: 872.3060  
Regulation Name: Alloy, Gold Based, For Clinical Use  
Regulatory Class: II  
Product Code: EJT  
Dated: October 18, 2001  
Received: October October 22, 2001

Dear Mr. Polzer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4618. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Timothy A. Ulatowski  
Director

Division of Dental, Infection Control  
and General Hospital Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510(k) Number (if known): K 012839

Device Name: Auropal 1

**Indications for Use:**

Auropal 1 is a gold-silver casting alloy that can be used by dental technicians to fabricate dental appliances for patients.

It is intended for manufacturing

- Inlays/Onlays
- Partial crowns
- Crowns
- Short span bridges
- Long span bridges
- Removable partials

and can be used for

- Telescopic and milling work

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Prescription Use ☒  
(Per 21 CFR 801.109)

OR

Over-the-Counter Use ☐

(Optional Format 1-2-96)

Angela Blackwell for M/S Runner  
(Device Sign-Off)  
Division of Dental, Infection Control,  
General Hospital Devices  
Number K012839